

<u>SL.NO</u>	<u>Material code</u>	<u>Material Description</u>	<u>Unit code</u>	<u>Req Qty (No.s)</u>
<u>1</u>		<p>Supply of Transformer oil BDV Test Kit:</p> <p>Specifications: Input: 230VAC, 1-Phase, 50Hz, output: 0-100KV, Center tap earthed capacity: 1KVA, Indication: Mains "on" & "OFF" LED lamps, Metering: 96Sq.mm Size analog / Digital voltmeter to monitor output voltage, Protection: Electronic, fast action over current tripping device sensing HT current directly, Motorisation: The unit will be supplied with the rate of rise of voltage @ 2KV/Second and in case of motorization failure, then can be operated manually, (Duty: int, Ref: SL.NO: SE/60/10-11). IS: 6792, IEC-156</p> <p>Safety features: 1) Zero inter-locking arrangement 2) 2) A transparent hood (cover) on test cup with door interlocking micro switch. 3) Resin encapsulated HT Transformer 4) The test cell is made from transparent high impact acrylic sheet which enables the operator to observe the flash over of the oil under test. construction: it is in portable air cooled sheet metal casting with powder coating consists of control cubical and H.T. transformer. 5) Magnetic stirrer facility shall be given. 6) Digital timer .</p> <p>Accessories: Go & NO Go Gauge, Test cell, Instruction manual.</p> <p>Make: Metravi/ Equivalent in udeyraj/cbs technologies /testing instruments/ neutonics /suyog/ BHEL/ Kirloshker / NMS/Zeal/Kalpa/Kappa.</p>	<u>01</u>	<u>01</u>

Pgs
22/12/18