

:

URANIUM CORPORATION OF INDIA LIMITED (UCIL)

(A Govt. of India Enterprise)

(CIN: U 12000 JH 1967 GOI 000806)

At: Tummalapalle Mine, (Near AMD Camp), PO: Mabbuchintalapalle
Vemula Mandal, District: YSR (Kadapa), Andhra Pradesh, Pin code – 516349
Tele fax: 08588-294424



Expression of Interest (EOI) in *ONE PART* is invited in prescribed format from interested Laboratories/ Radiological/ Diagnostic/ Imaging Centers located in Kadapa District to empanelled with UCIL Tummalapalle Unit for providing service to our employees and their dependents on credit basis under the CGHS Rates.

EOI No: UCIL/TMPL/P&A/EOI-01/2026

Date: 05/01/2026

Last date for receiving EOI in the Personnel Section of Tummalapalle Unit: 28.01.2026(up to 5PM)

DETAILS OF THE EMPANELMENT:

- A. Expression of Interest in one part is invited in prescribed format from interested Laboratories/ Radiological/ Diagnostic/ Imaging Centres' located in **Kadapa District** to empanelled with UCIL Tummalapalle Unit for providing service on credit basis for employees of Tummalapalle Unit. Under the CGHS Rates.
- B. The tenure of Agreement will be normally for a period of 03 years with two year further extension subject to mutual consent of both the parties. Laboratories/ Diagnostic/ Radiological/ Imaging Centers' having the capacity to cater to the medical needs of (i) about 600 departmental employees and their dependent family members; and (ii) approximately 150 employees of the agencies engaged by UCIL and their dependent family members; in accordance with their entitlements and their cases being individually referred to by Medical Officer of UCIL.

2. CATEGORIES OF HEALTH CARE CENTRES': The following categories of Health Care Centres' will be empanelled: Laboratories/ Diagnostic Centres'/ Radiological/ Imaging Centres'

3. GENERAL PRE-QUALIFICATION CRITERIA FOR EMPANELMENT: The request for empanelment by the Laboratory/ Diagnostic/Radiological/ Imaging Centre should be accompanied by the following details/ documents:

- i. Acceptance of treatment charges as applicable and approved from time to time by Central Govt. Health Scheme (CGHS) for respective cities.
- ii. Name of Centre with address including contact number, Fax number & e-Mail address..
- iii. Proof of the Centre being functional for last three years. Copy of the audited balance sheet, profit & loss account for the last three financial year, ending FY 2024-25.(Main documents/ summary sheet) shall be submitted by the concerned bidder.
- iv. Lists of investigation facilities available.
- v. Notarized copy of valid registration of the Lab/ any third-party certificate.
- vi. Registration under PNDT Act, if Ultrasonography facility is available.
- vii. Compliance with all statutory requirements including Bio Medical Waste (BMW) certificate from respective State Govt. Pollution Control Board/AERB/ Other Statutory Agency
- viii. PAN Number and GST Registration No. (if applicable)
- ix. Documents regarding installation of fire safety mechanism.
- x. 24 hours emergency services & facility for creation/ delivery of reports at the earliest, preferably on-line as well.

- xi. Average annual financial turn over during the last 3 years, ending 31st March of the previous financial year (2024-25), should be at least 27 Lakh.**
- xii. Provision for adequate sitting area, Lab Area and Imaging centre**
- xiii. They should have Fire and Safety measures in their centres.**
- xiv. Hospital/ Laboratory/ Diagnostic/Radiological/ Imaging Centre empanelled earlier with UCIL either in CGHS or below CGHS rate, if apply, than their bidding shall be out rightly rejected & shall not be considered for further evaluation.**
- xv. The received bids same contact number of any submitted documents shall be out rightly rejected & shall not be considered for further evaluation.**

4. SPECIFIC QUALIFICATION CRITERIA:

A. LABORATORIES/ DIAGNOSTIC CENTRES:

- i. Availability of Qualified doctors in respective fields like MD (Path), MD (Microbiology)/ DCP etc (or) tie up with professional institute or doctors for the same.**
- ii. Availability of properly trained Technician working under guidance of qualified doctors like MD (Path), MD (Micro)/ DCP and/or having DMLT diploma/ degree (or)tie up with properly trained Technician working under guidance of qualified doctors like MD (Path), MD (Micro)/ DCP and/or having DMLT diploma/ degree.**
- iii. Availability of Auto-analyzer to conduct the blood biochemistry, microbiology, sputum, stool, urine etc. for examination/ investigation on referral would be preferable.**
- iv. Facilities for Routine microbiological investigation like culture and sensitivity.**
- v. Facilities/tests which are not available in the region/concerned centre, may be taken from National NABL standard like Thyrocare, Ranbaxy Laboratories Ltd., SRL Diagnostics/Dr Lal PathLabs etc., however payment will be reimbursed to the concerned centre on the basis of applicable/prevailing CGHS rates.**

B. RADIOLOGICAL/ IMAGING CENTRES:

B1.FOR MRI CENTRES:

- i. MRI Machine with magnet strength of 1.5 Tesla or more**
- ii. Availability of Radiographer (technician) having requisite degree/ diploma (B.Sc. in Radiography) (or) tie-up with Radiographer (Technician) having requisite degree/ diploma (B.Sc. in Radiography)**

- iii. Reports analysis and authentication by Medical officers having requisite degree **or tie-up** with Medical officers having requisite degree for Reports analysis and authentication.

B2.FOR CT SCAN CENTRES:

- i. Facility for whole Body CT Scanner with scan cycle of less than one second (sub-second).
- ii. Availability of Technician having DMLT diploma and Radiographer (technician) having requisite degree/ diploma **(or) tie-up with** Technician having DMLT diploma and Radiographer (technician) having requisite degree/ diploma.
- iii. Reports analysis and authentication by Medical officers having requisite degree **(or) tie-up** with Medical officers having requisite degree for the same.

B3.FOR X RAY CENTRES':

- i. X- Ray machine with minimum current rating of 500 MA with image intensifier TV System.
- ii. Portable X-Ray machine with minimum current rating of 60 MA
- vi. Availability of Technician having requisite Degree/ Diploma **(or) tie-up** with Technician having requisite Degree/ Diploma.
- vii. Reports analysis and authentication by Medical officers having DMRD/ MD (Radiology) degree **(or) tie-up with** Medical officers having DMRD/ MD (Radiology) for Reports analysis and authentication.

B4.FOR USG/ COLOUR DOPPLER CENTRE/ ECHO:

- i. Standard equipment having convex, sector, linear probes of frequency ranging from 3.5 to 10 MHz with minimum three probes and provision/ facilities of trans vaginal/ trans rectal probes.
- ii. Registration under PNNDT Act.
- iii. Reports analysis and authentication by Medical officers having requisite degree **(or) tie-up** with Medical officers having requisite degree for Reports analysis and authentication.

6. PROCEDURE FOR EMPANELMENT:

- i. Offers of EOI (Expression of Interest) will be invited through **open advertisement in local newspapers and publication in UCIL Website.**
- ii. EOI with all the supporting documents shall be submitted through Courier /Speed Post in sealed envelope addressed to Manager (Pers.), UCIL, Tummalapalle Mine, PO: Mabbuchintalapalle, Vemula Mandal, Dist: YSR (Kadapa), AP – 516349 super-scribed with (i) ‘EOI No.’, (ii) ‘Name of the Work’, and (iii) ‘Closing date for submission of EOI’

so as to reach him on or before the due date and time fixed for the opening of the bids failing which the bid of the concerned party shall be rejected.

- iii. After receiving the EOIs, proper scrutiny will be done by an ***Empanelment /Technical Evaluation Committee.***
- iv. The Committee may cancel/ reject any/ all EOIs without assigning any reasons thereof.
- v. This contract doesn't have provision of giving sub contract to other party or firm.
- vi. Tenders not fulfilling the specified technical specifications will be rejected.
- vii. The Empanelment /Technical Evaluation Committee constituted by UCIL will evaluate all the applications (**received though only offline mode**) and screen those hospitals for further process of empanelment on the basis General and Specific Pre-qualification Criteria, as applicable and as the case may be and give recommendations for empanelment.

7. For treatments/ tests/ any other aspects under the Laboratories/ Diagnostic Centres'/ Radiological/ Imaging Centres' which were not described in prevailing/applicable CGHS Rate, Then, the party/ concerned Centre has to provide minimum 10% discount on their Maximum Retail Price (MRP) notified by the Laboratories/ Diagnostic Centres'/ Radiological/ Imaging Centres.
8. All technically qualified Laboratories/ Diagnostic Centres'/ Radiological/ Imaging Centres' will be considered for empanelment.
9. Conditional tender/bid will not consider for Technical evaluation and bid of the concerned party will be rejected.
10. Initially all received application will be evaluate on the basis of General Pre-qualification Criteria. Bids qualify for General Pre-qualification Criteria will process further evaluation on the basis of specific qualification criteria.
11. **Example:** One Bidder has qualify **Specific Qualification Criteria** only in Laboratories/ Diagnostic Centre after qualifying in General Pre-qualification Criteria, than the ***Empanelment /Technical Evaluation Committee will recommend the concerned bid to empanelled for*** Laboratories/ Diagnostic Centre.
12. **Example:** One Bidder has qualify **Specific Qualification Criteria** in any Sub-Part(B1 to B4) of Radiological/ Imaging Centres' after qualify in General Pre-qualification Criteria, than the ***Empanelment /Technical Evaluation Committee will recommend the concerned bid to empanelled for*** Radiological/ Imaging Centre.
13. **Example:** One Bidder has qualify **Specific Qualification Criteria** in Laboratories/ Diagnostic and any Sub-Part(B1 to B4) of Radiological/ Imaging Centres' after qualify in General Pre-qualification Criteria, than the ***Empanelment /Technical Evaluation Committee will***

recommend the concerned bid to empanelled for Laboratories/ Diagnostic and Radiological/ Imaging Centre.

14. Period of Empanelment: The empanelment will be initially for a period of three years and further extended for two years subject to satisfactory services rendered. Performance of the Centre will be liable to be reviewed any time. If services are not found satisfactory, the empanelment will immediately be cancelled. However, if services of the centre are found to be satisfactory, further extension of empanelment will be done with the mutual consent of both the parties.

15. PQC ADDITIONAL DOCUMENTS:

- A. The eligibility is to be decided strictly based on documents submitted at the time of receipt of EOI. But there is no bar to seek clarification or authentication of submitted documents. Also If desired, UCIL may ask the bidders to submit the additional documents relevant to eligibility in the EOI for ensuring the genuineness of the bidder. However, in case of poor response, with a view to increase the competition, admission of additional documents to meet the PQC-Criteria's may be allowed subject to the condition that:-
- B. 'Poor Response' implies when less than three bids are found suitable (i.e technically qualified) on the basis of submitted eligible documents as per EOI. The additional document should not be issued subsequent to last date of receipt of tender as mentioned in the EOI. The opportunity of submission of additional documents will be given to all the bidders.

16. Criteria for De-Empanelment: Empanelment will liable to be terminated for any of the reasons enumerated below, which are representative but not exhaustive:

- A.** Written unwillingness to provide service by the concerned centre with minimum five months notice period or by UCIL with same period of notice.
- B.** The number of visits of patients to a Diagnostic Centres being considered as inadequate/ meagre.
- C.** Owing to malpractice/ misconduct or rendering unsatisfactory services.
- D.** Legal conviction by any State/ Central Govt. Authority/ Court against the Diagnostic centre.
- E.** Poor quality of reports.

12. Legal Jurisdiction

In case of any dispute arising out of terms of empanelment of Laboratory/ Diagnostic/Radiological/ Imaging Centre legal jurisdiction will be limited to UCIL Headquarter.

ANNEXURE -I

**APPLICATION FORM FOR EMPANELMENT OF LABORATORIES/
DIAGNOSTIC /RADIOLOGICAL/ IMAGING CENTRE**

1. NAME OF THE CITY WHERE LABORATORY/ DIAGNOSTIC CENTRE/ **RADIOLOGICAL CENTRE/ IMAGING CENTRE** IS LOCATED

2. NAME OF LABORATORY/ DIAGNOSTIC CENTRE/ **RADIOLOGICAL CENTRE/ IMAGING CENTRE**

3. ADDRESS OF LABORATORY/ DIAGNOSTIC CENTRE/ **RADIOLOGICAL CENTRE/ IMAGING CENTRE**

4. CONTACT DETAILS:

Landline Telephone/Mobile Number		
Email		
Website Address		

5.	Whether NABL/ NABH recommended?	Yes	No
6.	Whether QCI recommended? (Wherever applicable)	Yes	No

• Space:	Yes	No
• Equipment:	Yes	No
• Manpower with designated qualification	Yes	No
• Quality control measures	Yes	No
• General requirement for pathological diagnostic centres	Yes	No

7. If applying for empanelment of Laboratories/ Diagnostic Centres (availability of facilities to be mentioned)

• Space:	Yes	No
• Equipment:	Yes	No
• Manpower with designated qualification	Yes	No
• Quality control measures	Yes	No
• General requirement for pathological diagnostic centres	Yes	No

8. If applying for empanelment **Radiological/ Imaging Centres** (availability of facilities to be mentioned, if yes)

9. Total Turnover during last three financial years

(Certificate from Chartered Accountant to be enclosed)

10. Any bidder can apply Laboratories/ Diagnostic Centres and **Radiological/ Imaging Centres, together or can apply Laboratories/ Diagnostic Centres or Radiological/ Imaging Centres, separately.**

SIGNATURE OF APPLICANT OR AUTHORIZED AGENT

ANNEXURE II

CERTIFICATE OF UNDERTAKING

1. It is certified that the particulars given are correct and eligibility criteria are satisfied.
2. That Laboratory/ Diagnostics/ Radiological/ Imaging Centre shall not charge beneficiaries of UCIL India higher than the CGHS notified rates.
3. That the facility/ procedure/ investigation for which rates have been mentioned are actually available in the Centre. Facilities/tests which are not available in the region/concerned centre, may be taken from National NABL standard like Thyrocare, Ranbaxy Laboratories Ltd., SRL Diagnostics/Dr Lal PathLabs etc., however payment will be reimbursed to the concerned centre on the basis of applicable/prevailing CGHS rates.
4. That if any information is found to be fake, Laboratory/ Diagnostic/ Radiological / Imaging Centre would be liable to be de-empanelled.
5. That Laboratory/ Diagnostic/ Radiological/ Imaging Centre has not been derecognized by CGHS or any State Government or other Organizations in past.
6. That no investigation by Central Government/ State Government or any statutory Investigating Agency is pending or contemplated against the Laboratory/ Diagnostic/ Radiological/ Imaging Centre at present.
7. That I/ We agree with the terms and conditions prescribed in the EOI document.
8. That Laboratory/ Diagnostic/ Radiological/Imaging Centre agrees to maintain Electronic Medical records and EHR as per the standards approved by Ministry of Health & Family Welfare.
9. That I/We are not under suspension at present nor have been blacklisted by any PSU/ Government Department/ Financial Organization/ Court.
10. I /We hereby certify that I/ We have read the entire terms and conditions of the Empanelment documents including Annexure(s), Schedules (s) etc. which is a part of the Contract Agreement and I/ We shall abide by the terms/ conditions/ clauses contained therein.

SIGNATURE OF APPLICANT OR AUTHORIZED AGENT

ANNEXURE III

DOCUMENTS (WHERE IS APPLICABLE) FOR GENERAL PRE-QUALIFICATION CRITERIA FOR EMPANELMENT.

Sl. No.	List of Documents Enclosed (wherever applicable)	Please specify
1	Acceptance of treatment charges as applicable and approved from time to time by Central Govt. Health Scheme (CGHS) for respective cities.	Yes/No
2	Name of Centre with address including contact number, Fax number & e-Mail address..	Yes/No
3	Proof of the Centre being functional for last three years. Copy of the audited balance sheet, profit & loss account for the last three financial year, ending FY 2024-25.(Main documents/ summary sheet) shall be submitted by the concerned bidder.	Yes/No
4	Lists of investigation facilities available.	Yes/No
5	Notarized copy of valid registration of the Lab/ any third-party certificate.	Yes/No
6	Registration under PNDT Act, if Ultrasonography facility is available.	Yes/No
7	Compliance with all statutory requirements including Bio Medical Waste (BMW) certificate from respective State Govt. Pollution Control Board/AERB/ Other Statutory Agency	Yes/No
8	PAN Number and GST Registration No. (if applicable)	Yes/No
9	Documents regarding installation of fire safety mechanism.	Yes/No
10	24 hours emergency services & facility for creation/ delivery of reports at the earliest, preferably on-line as well.	Yes/No
11	Average annual financial turn over during the last 3 years, ending 31st March of the previous financial year (2024-25), should be at least 27 Lakh.	Yes/No
12	Provision for adequate sitting area, Lab Area and Imaging center	Yes/No
13	They should have Fire and Safety measures in their centres.	Yes/No
14	Hospital/ Laboratory/ Diagnostic/Radiological/ Imaging Centre empanelled earlier with UCIL either in CGHS or below CGHS rate, if apply, than their bidding shall be out rightly rejected & shall not be considered for further evaluation.	Yes/No
15	The received bids same contact number of any submitted documents shall be out rightly rejected & shall not be considered for further evaluation.	Yes/No

SIGNATURE OF APPLICANT OR AUTHORIZED AGENT

ANNEXURE IV

DOCUMENTS (WHERE IS APPLICABLE) SPECIFIC QUALIFICATION CRITERIA FOR EMPANELMENT i.e LABORATORIES/ DIAGNOSTIC CENTRES

Sl. No.	List of Documents Enclosed (wherever applicable)	Please specify
1	Availability of Qualified doctors in respective fields like MD (Path), MD (Microbiology)/ DCP etc (or) tie up with professional institute or doctors for the same.	Yes/No
2	Availability of properly trained Technician working under guidance of qualified doctors like MD (Path), MD (Micro)/ DCP and/or having DMLT diploma/ degree (or)tie up with properly trained Technician working under guidance of qualified doctors like MD (Path), MD (Micro)/ DCP and/or having DMLT diploma/ degree.	Yes/No
3	Availability of Auto-analyzer to conduct the blood biochemistry, microbiology, sputum, stool, urine etc. for examination/ investigation on referral would be preferable.	Yes/No
4	Facilities for Routine microbiological investigation like culture and sensitivity.	Yes/No
5	Facilities/tests which are not available in the region/concerned centre, may be taken from National NABL standard like Thyrocare, Ranbaxy Laboratories Ltd., SRL Diagnostics/Dr Lal PathLabs etc., however payment will be reimbursed to the concerned centre on the basis of applicable/pervailing CGHS rates.	Yes/No

SIGNATURE OF APPLICANT OR AUTHORIZED AGENT

ANNEXURE V

DOCUMENTS (WHERE IS APPLICABLE) SPECIFIC QUALIFICATION CRITERIA FOR EMPANELMENT i.e RADIOLOGICAL/ IMAGING CENTRES(B1. FOR MRI CENTRES):

Sl. No.	List of Documents Enclosed (wherever applicable)	Please specify
1	MRI Machine with magnet strength of 1.5 Tesla or more	Yes/No
2	Availability of Radiographer (technician) having requisite degree/ diploma (B.Sc. in Radiography) (or) tie-up with Radiographer (Technician) having requisite degree/ diploma (B.Sc. in Radiography)	Yes/No
3	Reports analysis and authentication by Medical officers having requisite degree or tie-up with Medical officers having requisite degree for Reports analysis and authentication.	Yes/No

SIGNATURE OF APPLICANT OR AUTHORIZED AGENT

ANNEXURE VI

**DOCUMENTS (WHERE IS APPLICABLE) SPECIFIC QUALIFICATION CRITERIA FOR
EMPANELMENT i.e RADIOLOGICAL/ IMAGING CENTRES(B2: FOR CT SCAN CENTRES):**

Sl. No.	List of Documents Enclosed (wherever applicable)	Please specify
1	Facility for whole Body CT Scanner with scan cycle of less than one second (sub-second).	Yes/No
2	Availability of Technician having DMLT diploma and Radiographer (technician) having requisite degree/ diploma (or) tie-up with Technician having DMLT diploma and Radiographer (technician) having requisite degree/ diploma.	Yes/No
3	Reports analysis and authentication by Medical officers having requisite degree (or) tie-up with Medical officers having requisite degree for the same	Yes/No

SIGNATURE OF APPLICANT OR AUTHORIZED AGENT

:

ANNEXURE VII

**DOCUMENTS (WHERE IS APPLICABLE) SPECIFIC QUALIFICATION CRITERIA FOR
EMPANELMENT i.e RADIOLOGICAL/ IMAGING CENTRES(*B3: FOR X RAY CENTRES*):**

Sl. No.	List of Documents Enclosed (wherever applicable)	Please specify
1	X- Ray machine with minimum current rating of 500 MA with image intensifier TV System.	Yes/No
2	Portable X-Ray machine with minimum current rating of 60 MA	Yes/No
3	Availability of Technician having requisite Degree/ Diploma (or) tie-up with Technician having requisite Degree/ Diploma.	Yes/No
4	Reports analysis and authentication by Medical officers having DMRD/ MD (Radiology) degree (or) tie-up with Medical officers having DMRD/ MD (Radiology) for Reports analysis and authentication.	Yes/No

SIGNATURE OF APPLICANT OR AUTHORIZED AGENT

:

ANNEXURE VIII

DOCUMENTS (WHERE IS APPLICABLE) SPECIFIC QUALIFICATION CRITERIA FOR EMPANELMENT i.e RADIOLOGICAL/ IMAGING CENTRES(B4: FOR USG/ COLOUR DOPPLER CENTRE/ ECHO):

Sl. No.	List of Documents Enclosed (wherever applicable)	Please specify
1	Standard equipment having convex, sector, linear probes of frequency ranging from 3.5 to 10 MHz with minimum three probes and provision/ facilities of trans vaginal/ trans rectal probes.	Yes/No
2	Registration under PNDT Act	Yes/No
3	Reports analysis and authentication by Medical officers having requisite degree (or) tie-up with Medical officers having requisite degree for Reports analysis and authentication.	Yes/No

SIGNATURE OF APPLICANT OR AUTHORIZED AGENT